



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE

RECEIVED

By Carol Day at 10:09 am, Sep 01, 2015

Complete this report in duplicate at the time of the regular monthly preventative maintenance.
Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 092835	PRINTER SN 071.3383.132	DATE OF INSPECTION 08/28/2015
LOCATION OF INSTRUMENT (STREET AND CITY) 4525 Downs Drive St. Joseph, Mo 64507		TIME OF INSPECTION 10:40 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☒ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

☒ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

☒ PRINTER WORKING PROPERLY

☒ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☒ SIMULATOR SOLUTION

☐ COMPRESSED ETHANOL-GAS MIXTURE

☒ STANDARD SUPPLIER RepCo Marketing Inc. LOT # 15001 EXP. DATE 05/20/2017

☒ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN 3506 SIMULATOR EXP DATE 11/17/2015

☒ CALIBRATION CHECK -- (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .102

TEST 2 .101

TEST 3 .101

☒ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is operating within the limits set forth by the Missouri Department of Health and Senior Services. Solution provided by RepCo Marketing Inc. .100 solution, Lot# 15001 expires 05/20/2017.

INSPECTING OFFICER

SIGNATURE James Bench
TYPE II PERMIT NUMBER/EXPIRATION DATE
240032 02/11/2016

PRINT NAME
James Bench
TELEPHONE NUMBER
(816) 271-4438

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial not 092035
Version not 5000

TEST EPOCH 00072

Temp Date Time 210L
s/

Alt 2100L
08/28/15 18:41 .008
Calibration Check
21 08/28/15 18:41 .102

Subject Name

Subject I.D.

Operator Name: J.D.

Location
James Bush 2117

4525 DOWNS DR

St. Joseph MO 64507

AS IV Serial not 092035
Version not 5000

TEST EPOCH 00074

Temp Date Time 210L
s/

Alt 2100L
08/28/15 18:43 .006
Calibration Check
21 08/28/15 18:43 .101

Subject Name

Subject I.D.

Operator Name: J.D.

Location
James Bush 2117

4525 DOWNS DR

St. Joseph MO 64507

AS IV Serial not 092035
Version not 5000

TEST EPOCH 00075

Temp Date Time 210L
s/

Alt 2100L
08/28/15 18:46 .008
Calibration Check
21 08/28/15 18:46 .101

Subject Name

Subject I.D.

Operator Name: J.D.

Location
James Bush 2117

4525 DOWNS DR

St. Joseph MO 64507

AS IV Serial not 092035
Version not 5000

TEST EPOCH 00076

Temp Date Time 210L
s/

Alt 2100L
08/28/15 18:46 .006
Calibration Check
21 08/28/15 18:46 .102

Subject Name

Subject I.D.

Operator Name: J.D.

Location
James Bush 2117

4525 DOWNS DR

St. Joseph MO 64507

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 15001

EXPIRATION DATE: May 20, 2017 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JAMES E BENCH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/11/2014

NUMBER 240032


EXPIRES 2/11/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES


MD 580-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator BENCH, JAMES
Permit No 240032
Date Issued 2/11/2014 Date Expires 2/11/2016